



Product Service

Attestation of Conformity

No. N8 16 03 67174 004

Holder of Certificate: **SEOJIN INSTECH Co., Ltd.**

12, Sagimakgol-ro, 148beon-gil
Jungwon-gu
Seongnam-si, Gyeonggi-do 13207
REPUBLIC OF KOREA

Product: **Electrical equ. for measurement, control and laboratory use
(Displacement Type Level Switch)**

Model(s): **SMC**

Parameters:

Rated and principal characteristics:	125/250/480 V~ 5A; 125 V dc, 0.5 A; 250 V dc, 0.25 A
Rated frequency:	50/60 Hz
Protection class:	I
Installation Category:	III
Pollution degree:	2
Degree of protection against ingress of liquids:	IP65

Tested according to: EN 61010-1:2010

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2006/95/EC relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: CP5A01050216



Date, 2016-03-15 (Hyuk-Jun Kwon)

CE After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.